

# 校准证书

## CALIBRATION CERTIFICATE

证书编号:



第 1 页 共 3 页

Certificate No.

J202104238529-0001

Page

of

委托方

HK LEE HING INDUSTRY CO., LIMITED

Client

联络信息

See more at: <http://www.china-gauges.com/>

Contact Inf.

仪器名称

Test Probe for Measuring Surface Temperatures of IEC  
60335-2-6 Figure 104

Description

型号/规格

LX-T20

制造厂

HK LEE HING  
INDUSTRY CO.,  
LIMITED

Model/Type

Manufacturer

出厂编号

JLX2021G200996

管理号

-----

Serial No.

Asset No.

接收日期

2021年04月25日

校准日期

2021年05月10日

Receipt Date

Y M D

Cal. Date

Y M D

发布日期

2021年05月10日

Issued Date

Y M D

批准

Approved by

林谷江 (副总监)

审核

Inspected by

叶超男

校准

Calibrated by

蔡泽秋

证书专用章

(Stamp)

总部地址(Headquarters Add.): 广东省广州市黄埔大道西平云路163号

No.163.Pingyun Rd, West of HuangPu Ave.Guangzhou Guangdong China

实验室地址(Add.of the Lab): 广东省广州市黄埔大道西平云路163号

No.163.Pingyun Rd,West of HuangPu Ave.Guangzhou,Guangdong,China

联系电话(Tel.):400-602-0999

邮政编码(Postcode):510656

扫一扫验真伪

网站(Website):[http:// www.grgtest.com](http://www.grgtest.com)

电子邮件(E-mail):[grgtest@grgtest.com](mailto:grgtest@grgtest.com)

## 校准说明 DIRECTIONS OF CALIBRATION

证书编号: J202104238529-0001

第 2 页 共 3 页

### Certificate No.

Page of

- 1.本实验室的质量管理体系符合ISO/IEC 17025:2017标准的要求。  
(The quality system is in accordance with ISO/IEC 17025:2017.)
  - 2.本结果仅对本次校准样品有效。未经实验室批准,不得部分复制。如有疑问请在15个工作日内反馈。  
(The result is only valid for the calibrated sample.The certificate shall not be reproduced except in full,without the written approval of our laboratroy .please feedback to us within 15 days if you have any question.)
  - 3.本证书编号具有唯一性,后缀若带有“-Gx”的证书为替换证书,自发出后原证书即刻作废。  
(Each certificate has a unique number. The suffix of "-Gx" will be added to the number as a replacement of the old version. The original certificate will be officially invalid once the new certificate number is issued.)
  - 4.证书中最大允许误差、判定结果仅供参考,其中“P”代表“合格”,“F”代表“不合格”,“N/A”代表“不适用”。使用人员应结合实际测量需求,评估测量不确定度对符合性评定的影响。(MPE & judgement result in the datasheet is only for reference, "P" is "Pass", "F" is "Fail" and "N/A" is "Not Applicable".Whereas users should evaluate the effects of MU of calibration results on conformance assessment by actual measurement.)
  - 5.本次校准的技术依据及CNAS认可范围,超出范围的内容未被认可。详细认可范围请查看CNAS网站中注册编号为L0446的证书附件。(Reference document and accredited scope by CNAS for calibration, beyond which isn't accredited. Please see the attachment of certificate No.L0446 on CNAS website for details.)
- JJF 1409-2013 表面温度计校准规范(C.S. for the Surface Thermometers) 温度: (-40~400)°C

### 6. 本次校准使用的主要测量标准(Main Standards of Measurement Used in the Calibration.):

名称 Description	编号 Serial No.	证书号/有效期 Certificate No./ Due Date	溯源机构 Traceability Institute	技术特征 Technique Character
表面温度校验仪	A85257	J202101165607-0002 2022-02-17	广州广电计量检测股份有限公司	稳定性: $\pm 0.4^{\circ}\text{C}/10\text{min}$ ; 均匀性 $\leq 0.5^{\circ}\text{C}$
特稳携式校验仪	82002347	J202103296192-0012 2022-03-29	广州广电计量检测股份有限公司	0.01级

### 7. 校准地点、环境条件(Place and environmental conditions of the calibration):

地点 Place	广州计量二次仪表室 (2)	温度 Temperature	24.2 °C	相对湿度 Relative Humidity	52 %
-------------	---------------	-------------------	---------	---------------------------	------

8. 建议复校时间间隔: 1 年,送校单位也可按实际使用情况自主决定。  
Suggested calibration interval is 1 year or it can be altered depending on the actual usage of the user.



# 校准结果 RESULTS OF CALIBRATION

证书编号: J202104238529-0001

第 3 页 共 3 页

Certificate No.

Page of

一、外观: 符合要求

Appearance: Pass

二、各部分相互作用: 正常 ;

Interreaction: Normal ;

三、温度示值校准: 单位 (°C)

Temperature and heat display calibration:

校准点 Cal.Point	示值误差 Error	不确定度 $U(k=2)$
50	-3.4	0.8
75	-5.6	0.8
100	-7.4	0.8
125	-7.6	0.8

备注:

Notes:

结论(Conclusion): 按校准结果使用

1.本报告中的扩展不确定度是由标准不确定度乘以包含概率约为95%时的包含因子 $k$ 。

The expanded uncertainty is given in the report by the standard uncertainty multiplied by the probability of about 95% when the factor  $k$ .

2.依据(Reference document)

JJF 1059.1-2012 测量不确定度评定与表示

(JJF 1059.1-2012 Evaluation and Expression of Uncertainty in Measurement)

(以下空白)

(The below is blank)

广州广电计量检测股份有限公司  
印章